



1 Chlorhexidine hydrochloride

2 β -Glycyrrhetic acid

HPLC Conditions 1

Column CAPCELL PAK C₁₈ SG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase 10mmol/L SDS, H₂O / CH₃CN = 50 / 50, pH2.2 (H₃PO₄)
Flow rate 1.0 mL/min
Temperature 40 °C
Detect UV 248 nm

HPLC Conditions 2

Column Superiorex ODS : 4.6 mm i.d.×250 mm
Mobile phase 10mmol/L SDS, H₂O / CH₃CN = 50 / 50, pH2.2 (H₃PO₄)
Flow rate 1.0 mL/min
Temperature 40 °C
Detect UV 248 nm