



1 Cyclandelate

2 Dicyclohexyl phthalate(I.S.)

#### HPLC Conditions

Column	CAPCELL PAK C18 UG120 S5 : 4.6 mm i.d.×150 mm
Mobile phase	CH <sub>3</sub> CN / H <sub>2</sub> O = 80 / 20
Flow rate	0.7 mL/min
Temperature	25 °C
Detect	UV 228 nm
Inj. vol.	10 µL
Resolution	13.2