


## ロクロニウム

## Rocuronium

筋弛緩剤ロクロニウムとその拮抗薬スガマデクスを CAPCELL PAK C<sub>18</sub> MGIII S3 (2.0 mm i.d. x 75 mm) にて分離し, NQAD にて検出しました。両物質は UV 検出器では高感度が得られないため NQAD による検出が有効な手段です。

Rocuronium, one of the muscle relaxants, and sugammadex, its antagonist, were separated with CAPCELL PAK C<sub>18</sub> MGIII S3 (2.0 mm i.d. x 75 mm), and detected with Nano Quantity Analyte Detector (NQAD). Since a UV detector shows a low sensitivity for both compounds, the use of NQAD seems advantageous.



### 【HPLC Conditions】

Column	: CAPCELL PAK C <sub>18</sub> MGIII S3 ; 2.0 mm i.d. x 75 mm
Mobile phase	: A) 0.1 vol% HCOOH, B) CH <sub>3</sub> OH B 10 % (0 min) 90 % (10 min) 10 % (10.1 min) Gradient
Flow rate	: 200 µL/min
Temperature	: 40 °C
Detector	: NQAD 5600 (Evaporation 35 , Nebulizer 30 , Filter 2.5 sec)
Inj. vol.	: 5 µL
Sample dissolved in	: Each standard was dissolved together in water at 1 mg/mL, and diluted with mobile phase A to 10 µg/mL. 1 µg/mL = 1 ppm