



1 Indometacin
1.0 µg/mL

Conditions of Primary separation

Precolumn CAPCELL PAK MF Ph-1 : 4.6 mm i.d.×50 mm
Mobile phase 100mmol/L Phosphate buffer / CH₃CN = 90 / 10, pH6.9
Flow rate 1.0 mL/min
Detect UV
Inj. vol. 20 µL

Conditions of Final separation

Analytical column CAPCELL PAK C₁₈ UG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase 100mmol/L Phosphate buffer / CH₃CN = 50 / 50, pH2.3 (H₃PO₄)
Flow rate 1.0 mL/min
Detect UV 254 nm