



1 Benzyl chloride

1

2.5 µg/mL

#### Conditions of Primary separation and Focusing

Precolumn	CAPCELL PAK C18 MG S5 : 2.0 mm i.d.×35 mm
Mobile phase	H <sub>2</sub> O / CH <sub>3</sub> CN = 60 / 40 (v/v)
Flow rate	[valve position A] 200 µL/min [valve position B] 100 µL/min
Temperature	40 °C
Detect	UV
Inj. vol.	100 µL

#### Conditions of Final separation

Analytical column	CAPCELL PAK C18 MG S5 : 1.5 mm i.d.×150 mm
Mobile phase	H <sub>2</sub> O / CH <sub>3</sub> CN = 40 / 60 (v/v)
Flow rate	100 µL/min
Temperature	40 °C

Detect

UV 218 nm