



## 1 Anthocyanin

### HPLC Conditions

Column	CAPCELL PAK C18 UG120 S3 : 1.5 mm i.d.×150 mm
Mobile phase	H <sub>2</sub> O / CH <sub>3</sub> CN / 0.5mol/L Oxalic acid = 80 / 20 / 2
Flow rate	200 μL/min
Temperature	40 °C
Detect	UV 350 nm
Inj. vol.	1 μL