



1 Glycyrrhizic acid

HPLC Conditions

Column	CAPCELL PAK C18 UG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase	H <sub>2</sub> O / CH <sub>3</sub> CN = 62 / 38, pH2.5 (H <sub>3</sub> PO <sub>4</sub> )
Flow rate	1.0 mL/min
Temperature	40 °C
Detect	UV 254 nm