



1 Tranexamic acid

HPLC Conditions 1

Column CAPCELL PAK C18 UG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase 50mmol/L KH₂PO₄, 0.2vol% H₃PO₄
Flow rate 1.0 mL/min
Temperature 45 °C
Detect UV 210 nm
Inj. vol. 10 µL

HPLC Conditions 2

Column CAPCELL PAK SCX UG80 : 4.6 mm i.d.×250 mm
Mobile phase 0.1mol/L KH₂PO₄, H₂O / CH₃CN = 80 / 20, pH4.0 (H₃PO₄)
Flow rate 1.0 mL/min
Temperature 40 °C
Detect UV 210 nm
Inj. vol. 10 µL

