



1 Fructose

2 Glucose

3 Sucrose

#### HPLC Conditions

Column	CAPCELL PAK NH <sub>2</sub> UG80 S5 : 1.5 mm i.d.×250 mm
Reagent	1.0 mol/L LiOH
Flow rate of reagent	200 μL/min
Mobile phase	CH <sub>3</sub> CN / H <sub>2</sub> O = 80 / 20
Flow rate	100 μL/min
Temperature	40 °C
Detect	PAD
Inj. vol.	1.0 μL