



1 Fructose

2 Glucose

3 Sucrose

HPLC Conditions

Column CAPCELL PAK NH₂ UG80 S5 : 1.5 mm i.d.×250 mm

Reagent 1.0 mol/L LiOH

Flow rate of reagent 200 μL/min

Mobile phase CH₃CN / H₂O = 80 / 20

Flow rate 100 μL/min

Temperature 40 °C

Detect PAD

Inj. vol. 1 μL