



1 Berberine

#### HPLC Conditions

Column	CAPCELL PAK C <sub>18</sub> SG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase	CH <sub>3</sub> CN / H <sub>2</sub> O / CHCl <sub>3</sub> = 90 / 80
Flow rate	1.5 mL/min
Temperature	40 °C
Detect	UV 254 nm
Inj. vol.	10 μL