



1 Glycyrrhizic acid
0.01 g/mL

HPLC Conditions

Column	CAPCELL PAK C ₁₈ UG120 S5 : 4.6 mm i.d.×150 mm
Mobile phase	H ₂ O / CH ₃ CN / H ₃ PO ₄ = 650 / 350 / 1
Flow rate	1.0 mL/min
Temperature	R.T.
Detect	UV 254 nm
Inj. vol.	10 µL