



1 Barbitol sodium(I.S.)

2 Strychnine

HPLC Conditions

Column	CAPCELL PAK C18 UG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase	0.68w/v% KH ₂ PO ₄ / CH ₃ CN / (C ₂ H ₅) ₃ N = 90 / 10 / 2
Flow rate	0.8 mL/min
Temperature	R.T.
Detect	UV 210 nm
Inj. vol.	5 µL