



1 Glycyrrhizic acid

HPLC Conditions

Column CAPCELL PAK C18 MG S5 : 1.5 mm i.d.×250 mm

Mobile phase 2vol% CH<sub>3</sub>COOH / CH<sub>3</sub>CN = 3 / 2

Flow rate 100 μL/min

Temperature 40 °C

Detect UV 254 nm

Inj. vol. 2 μL