



1 Glycyrrhizic acid

HPLC Conditions

Column	CAPCELL PAK C18 MGII S5 : 4.6 mm i.d.×150 mm
Mobile phase	2vol% CH ₃ COOH / CH ₃ CN = 60 / 40
Flow rate	0.5 mL/min
Temperature	20 °C
Detect	UV 254 nm
Inj. vol.	10 µL