



1 Arbutin

HPLC Conditions

Column	CAPCELL PAK C ₁₈ MGII S5 : 4.6 mm i.d.×150 mm
Mobile phase	H ₂ O / CH ₃ OH / 0.1mol/L HCl = 94 / 5 / 1
Flow rate	0.9 mL/min
Temperature	20 °C
Detect	UV 280 nm
Inj. vol.	10 µL