



1 Capsaicin

2 Dihydrocapsaicin

HPLC Conditions

Column CAPCELL PAK Ph UG120 S5 : 4.6 mm i.d.×250 mm

Mobile phase 0.1vol% H₃PO₄ / CH₃CN = 3 / 2

Flow rate 0.8 mL/min

Temperature 30 °C

Detect UV 281 nm

Inj. vol. 20 µL