



1 Tryptic digest of casein

HPLC Conditions

Column	CAPCELL PAK C18 UG120 S5 : 1.0 mm i.d.×250 mm
Mobile phase	(A) 0.1vol% TFA, H ₂ O (B) 0.1vol% TFA, CH ₃ CN B 10% - 50% (45min) Gradient
Flow rate	70 μ L/min
Temperature	35 $^{\circ}$ C
Detect	UV 210 nm
Inj. vol.	1.0 μ L