



1 Lysozyme chloride

1

500 µg/L

HPLC Conditions

Column CAPCELL PAK C₈ SG300 S5 : 4.6 mm i.d.×150 mm

Mobile phase 0.1vol% TFA, CH₃CN / H₂O

Flow rate 1.5 mL/min

Detect UV 214 nm

Inj. vol. 10 µL