



## 1 Tryptic digest of casein

### HPLC Conditions

Column	CAPCELL PAK C18 MG S3 : 0.3 mm i.d.×150 mm
Mobile phase	(A) 0.05vol% TFA, H <sub>2</sub> O (B) 0.05vol% TFA, CH <sub>3</sub> CN B 10% (0min) - 50% (45min) - 50% (60min) Gradient
Flow rate	5 µL/min
Detect	UV 210 nm
Inj. vol.	60 nL