



1 Oxine-copper
0.02 ppm

HPLC Conditions

Column	CAPCELL PAK C18 SG120 S5 : 4.6 mm i.d.×250 mm
Mobile phase	{ 1mmol/L EDTA, 2vol% H ₃ PO ₄ } / CH ₃ CN = 96 / 4
Flow rate	1.0 mL/min
Temperature	40 °C
Detect	UV 250 nm
Inj. vol.	20 μL