



1 Formalin
1 1 μ L/mL

HPLC Conditions

Column	CAPCELL PAK C ₁₈ AG120 S5 : 4.6 mm i.d.×100 mm
Mobile phase	6mmol/L NaH ₂ PO ₄ , 0.1vol% H ₃ PO ₄
Flow rate	0.8 mL/min
Detect	VIS 412 nm*
Inj. vol.	10 μ L