



1 Mandelic acid

HPLC Conditions

Column Ceramospher Chiral RU-1 : 4.6 mm i.d.×250 mm

Mobile phase *n*-Hexane / C<sub>2</sub>H<sub>5</sub>OH / CH<sub>3</sub>COOH = 90 / 10 / 1

Flow rate 1.0 mL/min

Temperature 50 °C

Detect UV 254 nm