



## 【HPLC Conditions】

Column : CAPCELL PAK C<sub>18</sub> MGII S5 ; 4.6 mm i.d. x 150 mm  
 Mobile phase : A) 0.1 vol% HCOOH, B) 0.1 vol% HCOOH in CH<sub>3</sub>CN  
 B 30 % (0 min) -> 30 % (3 min) -> 70 % (10 min) -> 30 % (10 min)  
 -> 30 % (15 min) Gradient  
 Flow rate : 1 mL/min  
 Temperature : 40 °C  
 Detection : UV 260 nm  
 Inj. vol. : 5 μL  
 ※ 1 μg/mL = 1 ppm

## 【References】

1) Bunseki 2004, 9, The Japan Society for Analytical Chemistry