



1	ミオグロビンのトリプシン分解物
	Tryptic digest of myoglobin

HPLC Conditions	
Column	CAPCELL PAK C18 MG S3 : 0.3 mm i.d. × 150 mm
Mobile phase	(A) 0.05vol% HCOOH, H ₂ O (B) 0.05vol% HCOOH, CH ₃ CN B 10% (0min) 50% (60min) Gradient
Flow rate	5 μL/min
Detect	ESI-MS